

COPY 2

Morton

* granted extension
performed

3-24-04

VOC Emissions from Hemlock Lumber

Report to

Hampton Affiliates, Cowlitz Division
P.O. Box 189
Randle, WA 98377

Report by

Michael R. Milota
Department of Wood Science and Engineering
Oregon State University
Corvallis, OR 97331

April 15, 2004

RECEIVED
APR 22 2004

SOUTHWEST CLEAN
AIR AGENCY

TABLE OF CONTENTS

	Page
I Results Summary	1
II Lumber Source and Handling	1
III Kiln Description and Operation	2
Humidity	2
Temperature	3
Schedules	3
Charge sequence	5
IV Sampling Systems and Methodologies	7
V Data Reduction and Treatment	7
Flow calculations	7
Moisture calculations	8
Total hydrocarbon calculations	8
VI Sampling Results	9
VII Quality Assurance	16
Leak checks	16
Calibration gases	16
VIII Discussion and Recommendations	16

APPENDICES

Appendix 1. Detailed sampling procedures	17
Appendix 2. Data in electronic form	26
Appendix 3. Samples of field data sheets	27
Appendix 4. Calibration data	85

LIST OF FIGURES

FIGURE 1. Schematic of kiln and sampling system	2
FIGURE 2. Dry- and wet-bulb temperatures during the cycle	4
FIGURE 3A. Schematic of heated filter box.	7
FIGURE 3B. Photo of VOC sampling system	7
FIGURE 4. Hydrocarbon concentration and vent rate versus time	11
FIGURE 5. Cumulative emissions and rate of emissions versus time	12
FIGURE 6. Moisture content versus time for the charge	13
FIGURE 7. Cumulative emissions versus moisture content of the charge	14

LIST OF TABLES

TABLE 1. Summary of results	1
TABLE 2. Drying schedule	3
TABLE 3. Estimated VOC release at different final moisture contents.	10
TABLE 4. Summary of sample runs	15

VOC Emissions from ^{Hemlock} Sitka spruce Lumber

I. Results Summary

Two charges each containing 73.3 board feet of 2x4 hemlock lumber were dried in a small-scale kiln at Oregon State University. The kiln dry- and wet-bulb temperatures were provided by Hampton Lumber. One schedule was at a conventional temperature, with a maximum temperature of 180°F (82°C) and a wet-bulb temperature of 150°F (65°C). The other was at a high temperature, with a maximum temperature of 215°F (103°C) and a wet-bulb temperature between 150°F (65°C) and 165°F (74°C). The air velocity was 750 feet per minute (3.8 m/s). The kiln was indirectly heated with steam. There was no humidification. Regulating the amount of air entering the kiln controlled venting and the humidity.

A JUM 3-200 total hydrocarbon analyzer was used to measure organic emissions following EPA Method 25A. It has been demonstrated through past studies (Lavery and Milota, 2000, Forest Products Journal, NCASI/Georgia-Pacific SEP project) that this method in this small-scale kiln gives results similar to a large-scale kiln. The data for the test is summarized in the Table 1.

TABLE 1. Summary of results.

Charge	Schedule	Initial MC	VOC ^a	Time ^b
		%	lb/mbf	hrs
1	Conv.	115.7	0.40	52.9
2	High	112.9	0.34	32.7

^a VOC value reported at 15% moisture content. This can be adjusted to a different final moisture contents using data in report.

^b to 15% moisture content

II. Lumber Source and Handling

Three charges of lumber were delivered to Oregon State University on March 16, 2004, two to be dried and one as a backup. The wood was wrapped in plastic at the mill to prevent predying and loss of organic compounds.

The wood for two charges was rewrapped in plastic and stored at 2°C until they could be dried. The first charge was put in the kiln immediately; however, there was a problem with the kiln control and it was aborted. The conventional charge described in this report was started on March 24 and the high-temperature charge on March 29.

III. Kiln Description and Operation

A schematic of the kiln is shown in Figure 1. The kiln box is approximately 4' by 4' by 4'. It is indirectly heated by steam. Four dry-bulb thermocouples and two wet-bulb thermocouples are located on the entering-air side of the load. The dry-bulb thermocouples are spaced in a grid. The two wet-bulb thermocouples are under a single sock at the center of the entering-air side of the load.

Humidity control

A 200 L/min MKS mass flow meter controlled and measured the amount of air entering the kiln. It was factory calibrated and checked using a bubble meter. The amount of air entering the kiln is based on the wet-bulb temperature - if it is above setpoint, the airflow is increased and if it is below setpoint the airflow is decreased. This is analogous to venting for a commercial kiln. A minimum of 4 L/min entered the kiln at all times, more than removed by the analyzer (< 2.6 L/min). Putting air into the kiln at a rate of 100 L/min causes the pressure in the kiln to be 60 to 130 Pa above ambient, depending on location in the kiln (high-pressure or low-pressure side). Thus, any fugitive leakage should be out of the kiln. Two additional flow meters can be manually set to provide additional airflow. The steam spray line is disabled, so no water vapor is added to the kiln atmosphere.




FIGURE 1. Schematic of kiln and sampling system.

Temperature

Temperature in the kiln is controlled by indirect steam heating. When the average of the four dry-bulb thermocouples is below setpoint, the steam pressure in the coil is increased. When it is above setpoint, steam flow to the coil is reduced.

Schedules

The drying schedule supplied by the mill is shown in Tables 2a and 2b. The values in Table 2b are base on the entering-air temperature. This represents the highest temperature the wood would experience in the kiln. The actual temperatures in the lab kiln are presented in Figure 2a and 2b.

TABLE 2a. Drying schedule for the conventional-temperature charge.

Step time, hours	Ramp time, hours	Run time, hours	Dry-bulb, °F	Wet-bulb, °F
0	0	0	80	70
12	12	12	165	150
4	16	16 to dry	180	150

TABLE 2b. Drying schedule for the high-temperature charge.

Step time, hours	Ramp time, hours	Run time, hours	Dry-bulb, °F	Wet-bulb, °F
0	0	0	155	120
12	12	12	200	175
4	4	16	210	170
8	0	24	215	160
16	0	40	215	150
10	0	50 to dry	215	140




FIGURE 2a. Dry- and wet-bulb temperatures during the conventional-temperature drying cycle.




FIGURE 2b. Dry- and wet-bulb temperatures during the high-temperature drying cycle.

Charge Sequence

After warming to room temperature, 2" were trimmed from each end of each board to give 44" samples. These were then weighed, placed in the kiln, and dried according to the schedules in Tables 2a and 2b. Sampling for hydrocarbon was done as described in section IV. At the end of drying the wood was weighed, oven dried, and reweighed so initial and final moisture contents could be determined by ASTM D4442.

IV. Sampling Systems and Methodologies

Figures 3a and 3b show the hydrocarbon sampling system. Unlike stack testing, all necessary equipment is mounted on the kiln and flows are controlled with valves and switches. The THC sample was drawn from the kiln directly into a dilution/filter box mounted on the side of the kiln. The box was heated to 125°C. It is assumed that the gas in the kiln is well-mixed and that the composition in the kiln near the exhaust is the same at the composition of the exhaust. The sample line from the box to the analyzer was heated to 133°C. The valve at the back of the analyzer was heated to 145°C. Heated dilution gas can be added to the hydrocarbon sample gas to lower the gas moisture content to the detector.

The fuel gas was hydrogen. The span gas was EPA Protocol 905 ppm propane in air, the mid-gas was certified 412 ppm propane. The zero gas was Grade 5 air. Detailed sampling procedures are in Appendix 1 and a summary is presented below.

Leak checks were conducted before and after the charge was dried. Valves are closed and all components from just behind the probe tip to the valve at the back of the analyzer are placed under a 18-20 inHg vacuum. Less than one inHg pressure change during two minutes is acceptable and this was met.

Total flow and sample flow to the analyzer were checked using an NIST-traceable flow meter. This was done at the beginning and end of each sampling interval. The meter was attached to the system near the probe tip within the heated box. The valves were repositioned so that the sample came from the flow meter rather than the kiln. Readings of flow were made with the dilution gas off. The flow readings were verified by observing the change in the analyzer reading for span gas before and after the dilution gas was turned on. The dilution ratio calculated based on the analyzer readings was within 1 to 2% of that determined by the flow meter. Dilution was used when the gas moisture content in the kiln was greater than 15%.

Calibration of the zero and span of the detector was done at the beginning of each run (about every three hours with one five- and one seven-hour interval each night). The calibration gas was introduced by setting the valves so the calibration gas entered the system near the probe tip at ambient pressure. The calibration was checked at the end of

each run with no adjustments made to the zero or span during the run. The span drift was always less than two percent of full scale for a run and generally less than one percent. The zero drift was minimal during entire drying cycle.




FIGURE 3A. Schematic of heated filter box with air dilution system, heated sample line, and analyzer. Sample enters heated box from back of drawing (box is attached to kiln).




FIGURE 3B. Photo of VOC sampling system showing heated sample box (with white insulation), toggle valves and flow meter for calibration gases (upper left), on/off valve for calibration gas (3 at upper center right), heated sample line to analyzer (green tube, lower left), valve for sample (2 at center), toggle valve to vacuum pump (near calibration gas valves), and vent/flowmeter valve (4 at upper right).

V Data Reduction and Treatment

The "FlowCalc" page in the files "Hampton, Kiln.XLS" in Appendix 2 show the calculations for each 3-minute interval during the charge. Column A is a reading number. Columns B and C are the clock and charge times, respectively. Columns D and E are the average dry- and wet-bulb temperatures. Column F is the vapor pressure at the wet-bulb temperature. The absolute humidity is shown in column G and the molal humidity in column H.

Flow calculations

The volumetric dry gas flow rate in column I (files "Hampton, Kiln.XLS" in Appendix 2) is the flowmeter reading adjusted for the meter calibrations and the molar humidity of the entering gas. This is in standard (at 0°C) liters per minute. In column J this has been converted to a mass flow rate in kg/min and in column K is the same information is expressed as a molal flow rate.

Moisture calculations

The water removal rate in g/min (column L) (files "Hampton, Kiln.XLS" in Appendix 2) is calculated from the humidity and the gas flow rate and the total water (column M) is an integration of column L over time.

The moisture content of the wood at each interval in the event (column N) was determined by reducing the MC of the wood based on the amount of water leaving the kiln during the previous 3-minute interval.

Total hydrocarbon calculations

The original total hydrocarbon analyzer reading is shown in column O (files "Hampton, Kiln.XLS" in Appendix 2). In column P this has been corrected to compensate for the range setting switch on the analyzer and scaling between the analyzer reading and the computer reading. Also in column P, the THA data between sampling runs has been adjusted to the average of the data during the 12-minute periods before and/or after the analyzer testing and calibration time. The dilution THA (column Q) is the corrected THA reading divided by the dilution ratio (from column Y). In column R we have the opportunity to compensate for the effect of moisture on the JUM detector. This was not done so column R equals column Q. Finally in column S, the hydrocarbon concentration is converted to a dry gas basis concentration.

In column T the hydrocarbon flow rate in g/min as carbon is calculated in a manner analogous to the water flow rate using the dry gas flow rate and the hydrocarbon concentration. Column U is the integral of column T over time, the cumulative hydrocarbon release up to that point in the schedule. Column V is the cumulative unit emissions, that is, column U divided by the oven-dry weight of the wood in the kiln.

Column X indicates the hydrocarbon sampling run and column Y is the dilution ratio during that run. The next two columns, Z and AA, are the cumulative dry gas and water during the kiln cycle. These are used to obtain the average gas moisture contents. The corrected wood moisture content, as discussed in section VI, is shown in column AC. The kiln air and analyzer air moisture contents are shown in columns AD and AE.

At the end (bottom) of the FlowCalc spreadsheet are summaries by run of the flow data for the total hydrocarbon run intervals.

The other pages in the files "Hampton, Kiln.XLS" are graphs of the data in the FlowCalc page. Moisture content and board weight data are in the files named "Hampton, Board.XLS."

VI. Sampling Results

The hydrocarbon emissions are summarized graphically here. All emission data is presented in detail in Appendix 2. The graphs show the full drying cycle; however, 15% moisture content was reached at approximately 33 and 52 hours, respectively, for the conventional- and high-temperature schedules.

Figures 4a and 4b show total hydrocarbon concentration and vent rate versus time. The vent rate is high first 10 hours, and then decreases. The high vent rate is due to the low wet-bulb temperature and low humidity. The concentration remains low while the vent rate is high, then increases to a maximum. The concentration then continues to increase for the conventional-temperature schedule and decreases for the high-temperature schedule. This difference is probably due to the venting differences which result from maintaining the respective drying schedule. Note that the left ordinate in Figure 4b is on a different scale than in 4a. During the higher temperature schedule, there is a greater concentration, but the venting is less and for a shorter duration (33 vs. 52 hours).

Figures 5a and 5b show the cumulative hydrocarbon emissions and the rate of emissions versus time. The cumulative emissions is the emissions up to any point in time in the schedule. The rate of emissions is how much is coming out per unit time. The maximum occurs at approximately 15 to 20 hours in each schedule.

Figures 6a and 6b show the wood moisture content versus time. The estimated moisture content should most accurately represent the MC-time relationship because the initial and final moisture contents match the oven-dry test. The initial moisture contents were over 110%. The final moisture content was 10.7% for the wood dried using the conventional-schedule and 7.2% for the wood dried using the high-temperature schedule (at 33 and 52 hours, respectively). We also opened the kiln and weighed the wood at approximately 60 hours and 37 hours. The moisture contents at these times were 13.3 and 11.3%.

Figures 7a and 7b show the cumulative hydrocarbon emissions versus moisture content. The emissions for drying to any moisture content can be read from this graph. If, for example, a mill wanted to dry to 18% instead of 15%, the total hydrocarbon emissions could be estimated at 0.38 and 0.32 for the conventional- and high-temperature schedules.

TABLE 3. Estimated VOC release at different final moisture contents (read from data file).

Moisture content	VOC release	
	Conventional	High
%	lb/mbf	lb/mbf
12	0.416	0.351
13	0.410	0.346
14	0.405	0.341
15	0.400	0.335
16	0.395	0.330
17	0.391	0.325
18	0.386	0.320




FIGURE 4a. Hydrocarbon concentration and vent rate versus time for the conventional-temperature charge.




FIGURE 4b. Hydrocarbon concentration and vent rate versus time for the high-temperature charge.




FIGURE 5a. Cumulative emissions and rate of emissions versus time for conventional-temperature drying.




FIGURE 5b. Cumulative emissions and rate of emissions versus time for high-temperature drying.




FIGURE 6a. Moisture content versus time for the conventional-temperature charge.




FIGURE 6b. Moisture content versus time for the high-temperature charge.




FIGURE 7a. Cumulative emissions versus moisture content of the conventional-temperature charge.




FIGURE 7b. Cumulative emissions versus moisture content of the high-temperature charge.

Table 4 shows the VOC results by run for the charges. The interval time periods shown in the table include the times between sampling and mass calculations are adjusted to account for these. Sampling occurred for approximately 95% of the drying time to 15% MC. Copies of all field sampling sheets, including dilution system and heated component data, in Appendix 3.

TABLE 4a. Summary of sample runs for the conventional-temperature charge.

Sample Run	Time hrs	Dry Flow Rate @68 l/min	Wet Flow Rate @68 l/min	THC wet conc ppmv	THC mass as C lbs/mbf	THC rate as C lb/hr/mbf	Average Wood MC %	Average Air MC %	Average Anal. MC %
1	3.00	219.6	225.7	7.3	0.014	0.0046	115.4	2.7	2.7
2	3.00	244.8	258.7	6.9	0.015	0.0049	113.7	5.4	5.4
3	4.80	244.3	272.2	12.2	0.044	0.0092	108.6	10.2	7.7
4	7.55	165.6	208.7	19.2	0.081	0.0107	94.0	20.7	15.5
5	3.10	180.6	238.5	19.9	0.040	0.0130	76.5	24.3	15.0
6	2.75	179.4	236.8	19.9	0.035	0.0129	65.3	24.2	15.0
7	3.14	150.1	198.3	20.0	0.034	0.0110	59.9	24.3	15.1
8	2.25	125.3	165.3	20.9	0.021	0.0095	47.5	24.2	15.0
9	5.10	102.0	134.5	21.0	0.039	0.0077	38.8	24.1	15.0
10	7.90	69.2	91.0	22.3	0.044	0.0056	27.6	24.0	14.6
11	3.46	44.5	58.5	23.6	0.013	0.0038	20.7	23.8	14.5
12	3.15	34.3	44.9	24.7	0.010	0.0030	18.2	23.6	14.4
13	3.70	27.2	35.8	25.8	0.009	0.0025	16.0	24.0	14.7
Sum	52.90				0.400				
Ave.		137.5	166.8	18.7		0.008			

TABLE 4b. Summary of sample runs for the high-temperature charge.

Sample Run	Time hrs	Dry Flow Rate @68 l/min	Wet Flow Rate @68 l/min	THC wet conc ppmv	THC mass as C lbs/mbf	THC rate as C lb/hr/mbf	Average Wood MC %	Average Air MC %	Average Anal. MC %
1	3.00	189.7	222.0	15.8	0.028	0.0092	110.0	14.6	14.6
2	3.00	262.7	315.1	12.6	0.033	0.0109	97.6	16.6	8.5
3	4.80	140.2	182.1	23.2	0.053	0.0110	81.9	23.0	11.5
4	7.55	88.0	130.5	35.9	0.096	0.0127	59.6	32.6	12.4
5	3.10	91.3	126.3	33.0	0.035	0.0114	40.4	27.7	7.7
6	2.75	86.4	115.1	32.0	0.028	0.0101	32.4	25.0	11.0
7	3.05	76.2	99.2	30.9	0.025	0.0081	28.9	23.2	10.2
8	2.25	66.1	84.7	32.7	0.017	0.0076	21.0	22.0	10.8
9	3.20	57.8	73.1	33.3	0.021	0.0067	17.0	20.9	10.3
Sum	32.70				0.335				
Ave.		117.6	149.8	27.7		0.0097			

VII. Quality Assurance

Leak checks

Leak checks were performed on the VOC system before and after drying.

Calibration

Data for the calibration gases are given in Appendix 4. The mid gas was not named because the analyzer was within tolerance without naming. The calibration sheet for the flow meter is also included is also included in Appendix 4 as is the thermocouple calibration check.

VIII Discussion

There were no anomalies during the conventional- or high-temperature schedules that would affect the data. Data from the first charge, during which there was a data collection problem, is not reported here.

The VOC emission value was greater at the conventional temperature than at the high-temperature. This is inconsistent with what is generally expected; however, in the published literature one can occasionally find where this occurs. There is some variability in this type of data due to the wood. For "identical" replicates we have seen ranges in the past from 1.29 to 1.54 lb/mbf for ponderosa pine and 0.87 to 1.19 for lodgepole pine, 0.46 to 0.51 for Douglas-fir and 0.22 to 0.27 for white fir. Hemlock tends to be at least, if not more variable in wood properties than these other species. Therefore, it is possible that the wood in the kiln varied enough to give these results. We have observed values for hemlock during past studies from 0.07 to about 0.5 lb/mbf. We have seen the emissions from hemlock wood from the same source vary by a factor of two. Therefore, the results are not surprising.

Appendix 1. Detailed Sampling Procedures

INSTRUCTIONS FOR CHECKS OF EMISSIONS KILN

Purpose: Ensure kiln is operating correctly

Clock time: Record from computer

Run time: Record from computer. Check the box if the computer screen being refreshed and time is advancing.

Box temperature: Read from metal electrical box under desk, left controller. The top and bottom numbers should be similar on the box should be similar, about 126 C..

Valve temperature: Read from metal electrical box under desk, right controller. The top and bottom numbers should be similar on the box should be similar, about 154 C..

Dry-bulb temperature: Read from computer screen. Compare to graph to be sure it's correct. If it's not within a degree or two of the chart, check again in a few minutes. During startup (the first 3 or so hours), it may not be able to track. If it's too high, the heat valve should be closed, too low and the heat valve should be open. If it does not appear to be working correctly, call Mike or Mark.

Wet-bulb temperature: Read from computer screen. Compare to graph to be sure it's correct.

If it is too low, it means that the kiln atmosphere is too dry. Check the flow meters. If Flow1 is about 10 L/min (its lower limit), make sure that Flow2 and Flow3 are turned off

If it's too high, then either the kiln atmosphere is too humid or the sock is not being wetted. If Flow 1 is near 200 L/min (its upper limit) add venting by opening Flow2 and/or Flow 3. The maximum for Flow2 is 50 L/min, if it reads over this value for several readings, reduce it to about 45 L/min. Don't change Flow3 often, rather set it and leave it for several hours if possible. Keep the Flow 3 reading constant by small adjustments. As Flow1 decreases or Flow2 turned down, there is more pressure behind Flow3 and the flow increased. Check for water in the wet-bulb reservoir (push the float down and make sure it's getting water).

Check both Wet-bulb1 and Wet-bulb2 and make sure they are reading about the same. If they differ by more than 2 C, call Mike or Mark.

If both wet-bulbs are reading the same as the dry-bulb, check the wet-bulb water.

If these procedures do not correct the wet-bulb temperature within 30 minutes, call Mike or Mark.

Line temperature: Read from gray box on wall above analyzer. It should read about 275°F.

Chiller temperature: Read the chiller temperature. It should be about -1°C.

Flow 1: Read from computer. The value of Flow1 changes depending on the wet-bulb. If Flow 1 is 10 L/min and the wet-bulb is too low, there's probably nothing we can do. If it's 200 L/min and the wet-bulb is too high, Flow2 and/or Flow3 can be opened. Flow2 and Flow3 should be adjusted so that Flow1 stays below 175 to 200 L/min.

Flow 2: Read from computer. The value of Flow2 is set by you. It will vary a little - as flow 1 goes down, flow 2 will go up. Do not set it to < 40 L/min if you think Flow1 is going to decrease or it will go off scale and not be read by the computer

Flow 3: Read from meter. The value of Flow3 is set by you. It will vary a little - as flow 1 goes down, flow 2 will go up. Be sure to clearly record this value and when you change it

Dilution flow: Read dilution flow meter. It should read the same setting as the red flag. Do not adjust. If significantly different, investigate.

F/M Flow: Read from rotometer. This should be about 400 to 500 cc/min.

Line vacuum: Read from the vacuum gauge. This should be about 20"Hg.

INSTRUCTIONS - FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER PRE-SAMPLE PROCEDURE

BACKGROUND INFORMATION

Get the dry- and wet-bulb temperatures from the kiln schedule or off the computer. Use the highest expected values for the run.

Read absolute humidity off the psychrometric chart or table.

Calculate or read from tables -

$$\text{Percent moisture} = 100 / [1 + 1 / 1.61 * \text{AbHum}]$$

$$\text{Target Dilution Ratio (TDR)} = 15 / \text{Percent Moisture}$$

Event = the name of the drying cycle.

Run = the number of the 3-hour interval.

Operator, that's you.

Date and time are now, as you start the data collection process.

AMBIENT DATA

Call 9-754-0081 and get altimeter setting.

Read the laboratory temperature from the thermometer.

ANALYZER CALIBRATION

Set valves so that 1, 2 = off; 3=on; 4=vent. This allows gas to flow out of the vents from the calibration tanks and shuts off all other sources. Only calibration gas should go through the detector.

Open the zero gas tank valve

- zero toggle switch up (on), others down (off)
- set flow to 3.5 L/min using regulator on tank
- wait for a stable reading (about 30 to 60 seconds)
- use the zero dial (pot) on THA to get a zero reading
- read the analyzer
- read computer
- note pot setting
- close valve on zero gas tank

Open span gas tank valve

- span toggle switch up (on), others down (off)
- set flow to 3.5 L/min using regulator on tank
- set analyzer to range 3
- wait for a stable reading (about 30 to 60 seconds)

use the span dial (pot) on THA to get a reading of 905 ppm
read the analyzer, record, for example, 9.05 or 900
read computer (should read about 905)
note pot setting

Open mid gas tank valve

mid toggle switch up (on), others down (off)
set flow to 3.5 L/min using regulator on tank
wait for a stable reading (about 30 to 60 seconds)
read analyzer (do not adjust pot settings), record, for example, 4.12 or 412
read computer (should about 412)
check for within tolerance
turn off mid gas
all toggle switches off

SET DILUTION FLOW BEFORE RUN

Set valves so that 1, 2, 3 = off; 4=meter. This allows gas to flow only from the meter to the detector.

Use the Gilibrator to take 4 readings of the total flow rate (TFR). This is the total flow drawn by the analyzer and should be about 2.6 L/min

Make sure the average does not include any "bad" readings

Record the average, L/min = cc/min / 1000

Write the Event, Run, and "Pre-TFR" on the Gilibrator printout.

Calculate the next two values -

Target dilution flow rate (TDFR) is the $TFR \times (1 - DR)$

Target sample flow rate (TSFR) is the $TFR \times DR$

Check that the sum of these is the Total Flow Rate

Set dilution flow

Set red pointer to desired dilution flow (on meter with valve 1)

Slowly open lower valve on dilution flow meter (1=on; 2, 3=off; 4=meter)

Use upper valve on dilution flow meter to adjust flow

Do not adjust this meter after this point

Read the meter that you just set and record the value

Use the Gilibrator to take 4 readings of the sample flow rate (SFR). This is the flow through the analyzer after dilution is set. It will vary, depending on the dilution setting.

Make sure the average does not include any "bad" readings

Record the average, L/min = cc/min / 1000

Write "Pre-SFR" on the Gilibrator printout.

CHECK DILUTION FLOW BEFORE RUN

Set valves so that 1, 3 = on; 2=off; 4=vent. This allows gas to flow out of the vent from the calibration tank and shuts off all other sources. Calibration gas and dilution air will go through the detector.

Open span gas tank valve

span toggle switch up (on), others down (off)

set flow to 3.5 L/min using regulator on tank

set analyzer to range 3

wait for a stable reading (about 30 to 60 seconds) record

turn off all calibration gas tank valves

all toggle switches off

Calculate the dilution ratio based on gas flow by dividing the Sample Flow Rate by the Total Flow Rate.

Calculate the dilution ratio based on span gas by dividing the Diluted span by the undiluted span.

If the Dilution ratios do not agree within 5% - DO NOT PROCEED****. Use $100 * (DR_{\text{Span}} - DR_{\text{Flow}}) / DR_{\text{Flow}}$ to calculate the % difference.

**** check calculations, check that values for ppm and flows make sense, remeasure everything. If it still does not agree, call Mike or Mark

START RUN

Set valve so that 1, 2, 5 = on; 3, 4=off; all calibration tank valves off

Record the start time. Use the computer clock for all times or set your watch to the computer time.

Make sure analyzer is on appropriate range, usually range 3, to keep THC reading on computer between 60 and 750.

Monitor system, as needed. Record system condition at least hourly.

End time should be no more than 3 hours from start time.

POST-SAMPLE PROCEDURE

AT END OF RUN

Record your name as the operator.

Event = the drying cycle. Run = the 3-hour interval.

Operator, that's you. Date and time are now, as you start the data collection process.

AMBIENT DATA

Call 9-754-0081 and get temperature and altimeter

Local pressure = (Altimeter - 0.23) x 3.3867

Read the laboratory temperature from the thermometer.

Fill out appropriate information on Pre-sample side of data sheet for next run. This will save time in between runs.

END TIME

Record computer time.

DO NOT adjust dilution gas yet.

CHECK DILUTION FLOW AFTER RUN

Set valves so that 1, 3 = on; 2=off; 4=vent. This allows gas to flow out of the vent from the calibration tank and shuts off all other sources. Calibration gas and dilution air will go through the detector.

Open span gas tank valve

span toggle switch up (on), others down (off)

set flow to 3.5 L/min using regulator on tank

wait for a stable reading (about 30 -60 seconds)

record

all toggle switches off

Sample flow rate. Set valves so that 1=on; 2, 3 = off; 4=meter. This allows gas to flow only from the meter and the dilution to the detector.

Use the Gilibrator to take 5 readings of the sample flow rate (SFR). This is the flow through the analyzer with dilution on.

Make sure the average does not include any "bad" readings

Record the average, L/min = cc/min / 1000

Write "Post-SFR" on the Gilibrator printout.

Read dilution flow meter
To calculate the L/min, divide scfh by 2.12
Turn off dilution flow meter using valve 1

Total flow rate. Set valves so that 1, 2, 3 = off; 4=meter. This allows gas to flow only from the meter to the detector.

Use the Gilibrator to take 5 readings of the total flow rate (TFR). This is the total flow drawn by the analyzer and should be about 2.6 L/min
Make sure the average does not include any "bad" readings
Record the average, L/min = cc/min / 1000
Write "Post-TFR" on the Gilibrator printout.

CHECK CALIBRATION OF ANALYZER

Set valves so that 1, 2 = off; 3=on; 4=vent. This allows gas to flow out of the vents from the calibration tanks and shuts off all other sources. Only calibration gas should go through the detector.

Span gas tank valve should be open

- span toggle switch up (on), others down (off)
- set flow to 3.5 L/min using regulator on tank
- set analyzer to range 4
- wait for a stable reading (about 30 -60 seconds)
- read analyzer (do not adjust pot settings), record, for example, 1.50 as 1500
- read computer (should read about 152 due to range 4 setting)
- note pot setting
- check for within tolerance - between 1483 and 1573

Open mid gas tank valve

- mid toggle switch up (on), others down (off)
- set flow to 3.5 L/min using regulator on tank
- set analyzer to range 3
- wait for a stable reading (about 30 -60 seconds)
- read analyzer (do not adjust pot settings), record, for example, 8.50 as 850
- read computer (should read same as analyzer)
- check for within tolerance

Open the zero gas tank valve

- zero toggle switch up (on), others down (off)
- set flow to 3.5 L/min using regulator on tank
- wait for a stable reading (about 30 -60 seconds)
- read analyzer (do not adjust pot settings)
- read computer

note pot setting

Calculate the dilution ratio based on gas flow by dividing the Sample Flow Rate by the Total Flow Rate.

Calculate the dilution ratio based on gas flow by dividing the Sample Flow Rate by the Total Flow Rate.

Calculate % difference as $100 * \{ \text{Absolute Value } (DR_{\text{Span}} - DR_{\text{Flow}}) \} / DR_{\text{Flow}}$

Record the time now as the end time for check.

Tear off the four sets of Gilibrator readings (Pre-TFR, Pre-SFR, Post-SFR, Post-TFR) and staple to paper with other records.

Start Pre-Sample procedure for next run.

Appendix 2. Data in Electronic Form

CD in
copy 1

Appendix 3. Field data sheets.

Charge: 2	Date	Time
Hampton Affiliates	Start: 3-24-04	11:06
Page: 2	End:	

Clock time	Run Start time	Run hrs	Run #	Temperatures							Flows					Line Vac. inHg	
				Box °C	Valve °C	Dry- bulb °C	Wet- bulb °C	Line °F	Anlz °C	Chiller °C	1 L/min	2 L/min	3 SCFH	✓	Dilution SCFM		F/M ml/min
11:40	0:02		6	125	144	83	67	274			200	0	0	2			
1:42	2:06		7	124	145	83	66	272			190	0	0	2			
2:58	3:22		8	126	145	82	66	274			200	0	0	2			
5:07	5:27		8	125	145	83	66	273			47	0	0	2			
9:44	10:09		9	125	145	83	66	274			4	0	0	2			
5:56	18:21		11	125	145	83	66	274			37	0	0	2			
7:11	1:06		11	125	145	83	66	274			4	0	0	2			
9:23	0:05		12	125	145	83	63	273			4	0	0	2			
11:16	0:10		12	124	145	83	64	273			200	0	0	2			
12:17	0:00		12	125	145	83	63	273			180	0	0	2			
1:14	0:59		13	125	144	83	66	273			23	0	0	2			
2:01	1:46		13	125	144	83	66	273			29	0	0	2			
3:10	2:55		13	125	145	82	66	273			19	0	0	2			
5:00	Stopped	Kiln	Stopped	125	145	82	MC	273			10 A	50					
10:16	0:07		15	125	145	82	50	273									

3-26-04

Charge: #2	Date	Time
Hampton Affiliates	Start: 3-24-04	11:06 AM
Page: 1	End:	

Clock		Run	Run	Run	Temperatures								Flows					Line			
time	time	hrs	time	✓	Box	Valve	Dry-bulb	✓	Wet-bulb	✓	Line	Anlz	Chiller	1	2	3	✓	Dilution	F/M	Vac.	
					°C	°C	°C		°C		°F	°C	°C	L/min	L/min	SCFH		SCFM	ml/min	inHg	
11:03	Pre			✓	125	144	—		—		273	180	—	—	—	—	—	—	—	—	
11:09	3:14 ^{SEC}				126	144	25	✓	20	✓	273	181	—	4	0	0	—	0	—	—	
11:35	0:29			✓	124	145	29	✓	25	✓	273	180	—	200	0	0	—	0	—	—	
1:28	2:22			✓	125	145	35		31		273	180	—	200	38	0	—	0	—	—	
2:04	2:58			✓	125	145	36		34		273	180	—	194	38	0	—	0	—	—	
2:47	3:41			✓	125	144	39		36		274	180	—	194	38	0	—	0	—	—	
5:06	6:01			✓	125	145	47		43		274	180	—	194	37	0	—	0	—	—	
9:47	10:41			✓	124	144	62		54		274	—	—	194	36	30	—	0	—	—	
10:13	11:04				125	145	62		55		274	—	—	196	29	0	—	13	—	—	
5:18	18:02				125	145	78		65		274	—	—	177	32	0	—	13+	—	—	
5:39	18:34				125	145	78		65		274	—	—	126	32	0	—	2.0	—	—	
7:42	20:36				125	145	80		65		274	—	—	153	0	0	—	2.0	—	—	
8:43	21:37				126	145	81		66		273	—	—	154	0	0	—	2.0	—	—	
10:06	23:00				125	144	82		66		274	—	—	183	6	0	—	2	—	—	
11:31	24:25				125	146	83		66		274	—	—	132	0	0	—	2	—	—	
Reset program					saved				first st of data												

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: ~ 100

Event (kiln charge): Hampton # 2

Wet-bulb temperature: ~ 90

Run (sample): 1

Absolute humidity: < 0.03

Operator: Milota

Percent moisture: < 4

Date: 3-24-04

Target Dilution Ratio (TDR): 0

Time now: 10:50a

H_2 press = 2250 psi

AMBIENT DATA

Altimeter setting: 29.99 inHg

Laboratory temperature: 24 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	0 (0)	—	does not apply	482
span	905 (1527)	—	does not apply	203
mid	413	—	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.575 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) L/min [TFR x (1 - DR)]

sample flow rate (TSFR) L/min [TFR x DR]

Set and read dilution meter: scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	—	—	—	—

START TIME: 11:06

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 3

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milora

Event (kiln charge): Hampton 1 2

Time now: 2:06

Run (sample): 1

AMBIENT DATA

Airport pressure: 29.99 inHg

Laboratory temperature: 24 °C

END TIME: 2:06

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}		

Sample flow rate (SFR) : _____ L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: _____ scfh _____ L/min [L/min = scfh*0.472]

Total flow rate (TFR): _____ L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): _____ [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	905	905	875 to 935	
mid	412	413	379 to 437	none
zero	0	1	-45 to +45	

Dilution ratio (DR_{Span}): _____ [Span_{Diluted} / Span]

Dilution ratio difference: _____ % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: _____

Comments: SPAN EPA Protocol 905 PPM } CH₄ in air
Mid Testifier 412 ppm } 38
Leak check @ 10:53 18.5 inHg VAC → 17.9 10:56

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: Lo

Event (kiln charge): Hampton #2

Wet-bulb temperature: Lo

Run (sample): 2

Absolute humidity: Lo

Operator: Milota

Percent moisture: <5%

Date: 3-24-04

Target Dilution Ratio (TDR): —

Time now: 2:00

AMBIENT DATA

Altimeter setting: 29.99 inHg

Laboratory temperature: 24 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>482</u>
span	<u>905</u> (1527)	<u>905</u>	does not apply	<u>203</u>
mid	<u>412</u>	<u>413</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.582 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: _____ scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): _____ L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}				

START TIME: 2:10

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 3

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milots

Event (kiln charge): Hampton # 2

Time now: 5:06 p

Run (sample): 2

AMBIENT DATA

Airport pressure: 29.99 inHg

Laboratory temperature: 24 °C

END TIME: 5:06

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}		

Sample flow rate (SFR) : L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>9023</u>	<u>903</u>	875 to 935	<u>48203</u>
mid	<u>413</u>	<u>414</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>482</u>

Dilution ratio (DR_{Span}): [Span_{Diluted} / Span]

Dilution ratio difference: % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check:

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: M. J. TA

Event (kiln charge): Hampton 1

Time now: 9:50

Run (sample): 43

AMBIENT DATA

Airport pressure: 29.97 inHg

Laboratory temperature: 25 °C

END TIME: 10:00

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}		

Sample flow rate (SFR) : _____ L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: _____ scfh _____ L/min [L/min = scfh*0.472]

Total flow rate (TFR): _____ L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): _____ [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	902	903	875 to 935	202
mid	412	413	379 to 437	none
zero	0	0.5	-45 to +45	482

Dilution ratio (DR_{Span}): _____ [Span_{Diluted} / Span]

Dilution ratio difference: _____ % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 10:02

Comments:

Analyzer in thermal T is spastic,
but analyzer is stable

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 40-50C
 Wet-bulb temperature: 35-40C
 Absolute humidity: 10-
 Percent moisture: 10-15
 Target Dilution Ratio (TDR): —

Event (kiln charge): Hampton 1
 Run (sample): 3
 Operator: Milota
 Date: 3-24-04
 Time now: 5:05

AMBIENT DATA

Altimeter setting: 29.99 inHg

Laboratory temperature: 24 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	903 (0)	1	does not apply	482
span	903 (1527)	903	does not apply	203
mid	413	414	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2,600 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) — L/min [TFR x (1 - DR)]

sample flow rate (TSFR) — L/min [TFR x DR]

Set and read dilution meter: — scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): — L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}				

START TIME: 5:10

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 3

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 60-70 °C

Event (kiln charge): Hampton # 2

Wet-bulb temperature: 53-60 °C

Run (sample): 4

Absolute humidity: ~ 0.1

Operator: Milota

Percent moisture: ~ 15

Date: 3-24-04

Target Dilution Ratio (TDR): 0.8

Time now: 9:45

AMBIENT DATA

Altimeter setting: 29.97 inHg

Laboratory temperature: 25 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	0 (0)	1	does not apply	482
span	905 (1527)	906	does not apply	204
mid	413	413	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.605 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 0.47 L/min [TFR x (1 - DR)]

sample flow rate (TSFR) 2.13 L/min [TFR x DR]

Set and read dilution meter: 13 0.996 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.945 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer 677 ppm	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>0.748</u>	<u>0.748</u>	<u>0.748</u>	<u>0.13</u>

START TIME: 10:10

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 3

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: M. Lota

Event (kiln charge): Hampton # 2

Time now: 5:19 a

Run (sample): 4

AMBIENT DATA

Airport pressure: 29.75 inHg

Laboratory temperature: 26 °C

END TIME: 5:19

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	666	666

Sample flow rate (SFR) : 1.965 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 1.33 scfh 0.628 L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2.598 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.756 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	891	892	875 to 935	204
mid	407	407	379 to 437	none
zero	0	1.5	-45 to +45	482

Dilution ratio (DR_{Span}): 0.747 [Span_{Diluted} / Span]

Dilution ratio difference: 1.2 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 5:26a

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 77

Event (kiln charge): Hampton #2

Wet-bulb temperature: 66

Run (sample): 5

Absolute humidity: 0.2

Operator: Milota

Percent moisture: 27

Date: 3-25-04

Target Dilution Ratio (TDR): 0.63

Time now: 5:27

AMBIENT DATA

Altimeter setting: 29.75 inHg

Laboratory temperature: 26 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>905</u>	does not apply	<u>210</u>
mid	<u>412</u>	<u>412</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.606 L/min

[1, 2, 3 = off; 4=meter]

Target

dilution flow rate (TDFR) 1.64 L/min

[TFR x (1 - DR)]

sample flow rate (TSFR) 0.964 L/min

[TFR x DR]

Set and read dilution meter: 2.1 scfh

[scfh = L/min * 2.12]

Sample flow rate (SFR): 1.611 L/min

[1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>559</u>	<u>0.617</u>	<u>0.618</u>	<u>0.16%</u>

START TIME: 5:33

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 3

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton #2

Time now: 8:33a 3-25-04

Run (sample): 5

AMBIENT DATA

Airport pressure: 29.70 inHg

Laboratory temperature: 25 °C

END TIME: 8:33

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>555</u>	<u>556</u>

Sample flow rate (SFR) : 1.607 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2.583 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.622

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>902</u>	<u>903</u>	875 to 935	<u>210</u>
mid	<u>412</u>	<u>413</u>	379 to 437	none
zero	<u>0</u>	<u>2</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}): 0.615

[Span_{Diluted} / Span]

Dilution ratio difference: 1.12 %

[100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 8:37

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82 Event (kiln charge): Hampton # 2
 Wet-bulb temperature: 66 Run (sample): 6
 Absolute humidity: 0.2 Operator: Milota
 Percent moisture: 26 Date: 3-25-04
 Target Dilution Ratio (TDR): 0.6 Time now: 8:33

AMBIENT DATA

Altimeter setting: 29.70 inHg Laboratory temperature: 25 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>905</u>	does not apply	<u>212</u>
mid	<u>413</u>	<u>412</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.589 L/min [1, 2, 3 = off; 4=meter]
 Target
 dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]
 sample flow rate (TSFR) ~~2~~ L/min [TFR x DR]
 Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]
 Sample flow rate (SFR): 1.606 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>560</u>	<u>0.619</u>	<u>0.620</u>	<u>0.16%</u>

START TIME: 842 [1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 3 [60 < computer reading < 750]

Hydrogen
2020
PSI

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton # 2

Time now: 11:38

Run (sample): 26

AMBIENT DATA

Airport pressure: 29.64 inHg

Laboratory temperature: 24 °C

8:42A

END TIME: 11:40

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	558	559

Sample flow rate (SFR) : 1.608 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2.586 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.622 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	901	902	875 to 935	212
mid	412	411	379 to 437	none
zero	0	1.5	-45 to +45	480

Dilution ratio (DR_{Span}): 0.618 [Span_{Diluted} / Span]

Dilution ratio difference: 0.55 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 11:44

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton 2

Wet-bulb temperature: 66

Run (sample): 7

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-25-04

Target Dilution Ratio (TDR): 0.6

Time now: 11:38

AMBIENT DATA

Altimeter setting: 29.64 inHg

Laboratory temperature: 24 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1.5</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>215</u>
mid	<u>414</u>	<u>415</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.587 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.612 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>561</u>	<u>0.612</u>	<u>0.623</u>	<u>0.52%</u> 1.8%

START TIME: 11:48

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 3 → 2

[60 < computer reading < 750]

@ 11:52
Vent or Meter open until 11:54
BAD DATA 11:48 → 11:54

H₂ at
2000 psi

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 1

Time now: 2:48

Run (sample): 7

AMBIENT DATA

Airport pressure: 29.70 inHg

Laboratory temperature: 24 °C

END TIME: 2:48

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>557</u>	<u>557</u>

Sample flow rate (SFR): 1,589 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2,578 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.618 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>905</u>	<u>905</u>	875 to 935	<u>215</u>
mid	<u>415</u>	<u>416</u>	379 to 437	none
zero	<u>0</u>	<u>0</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}): 0.616 [Span_{Diluted} / Span]

Dilution ratio difference: 0.32 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 2:51

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton # 2

Wet-bulb temperature: 66

Run (sample): 8

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-25-04

Target Dilution Ratio (TDR): 0.6

Time now: 2:38

AMBIENT DATA

Altimeter setting: 29.70 inHg

Laboratory temperature: 24 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0</u>	does not apply	<u>488</u>
span	<u>905</u> (1527)	<u>905</u>	does not apply	<u>215</u>
mid	<u>415</u>	<u>416</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.573 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.593 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>561</u>	<u>0.612</u>	<u>0.619</u>	<u>1.1 %</u>

START TIME: 2:54

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 1 2

Time now: 5:03 p

Run (sample): 8

AMBIENT DATA

Airport pressure: 29.76 inHg

Laboratory temperature: 24 °C

END TIME: 5:04

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>560</u>	<u>560</u>

Sample flow rate (SFR) : 1,589 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2,572 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.618

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>906</u>	<u>907</u>	875 to 935	<u>215</u>
mid	<u>414</u>	<u>416</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}): 0.618

[Span_{Diluted} / Span]

Dilution ratio difference: 0 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 5:08

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton # 2

Time now: 9:43

Run (sample): 9

AMBIENT DATA

Airport pressure: 29.82 inHg

Laboratory temperature: 23 °C

END TIME: 9:46 pm

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>562</u>	<u>562</u>

Sample flow rate (SFR) : 1,602 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2,580 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.621 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>909</u>	<u>909</u>	875 to 935	<u>214</u>
mid	<u>418</u>	<u>418</u>	379 to 437	none
zero	<u>0</u>	<u>0</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}): 0618 [Span_{Diluted} / Span]

Dilution ratio difference: ~~0.64~~ 0.64% [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 9:50

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton # 2

Wet-bulb temperature: 66

Run (sample): 9

Absolute humidity: 1

Operator: Milota

Percent moisture: 1

Date: 3-25

Target Dilution Ratio (TDR): 0.6

Time now: 5:03

AMBIENT DATA

Altimeter setting: 29.76 inHg

Laboratory temperature: 23 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>214</u>
mid	<u>415</u>	<u>414</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2,578 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1,578 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>559</u>	<u>0.618</u>	<u>0.612</u>	<u>0.9</u>

START TIME: 5:13

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton 1 2

Wet-bulb temperature: 66

Run (sample): 10

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-25-04

Target Dilution Ratio (TDR): 0.6

Time now: 9:43

AMBIENT DATA

Altimeter setting: 29.82 inHg

Laboratory temperature: 23 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>905</u>	does not apply	<u>211</u>
mid	<u>413</u>	<u>414</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.575 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.594 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>9153</u>	<u>0.611</u>	<u>0.619</u>	<u>1.13</u>

START TIME: 9:54

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 1 2

Time now: 5:39a 3-26-04

Run (sample): 10

AMBIENT DATA

Airport pressure: 539 inHg

Laboratory temperature: 23 °C

29.93

END TIME: 5:39

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>556</u>	<u>557</u>

Sample flow rate (SFR) : 1,586 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh 1 L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2,579 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}):

0.615

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>907</u>	<u>908</u>	875 to 935	<u>211</u>
mid	<u>415</u>	<u>416</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}):

0.614

[Span_{Diluted} / Span]

Dilution ratio difference:

0.16 %

[100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check:

5:48

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82
 Wet-bulb temperature: 66
 Absolute humidity: 5
 Percent moisture: 5
 Target Dilution Ratio (TDR): 0.6

Event (kiln charge): Hampton 1
 Run (sample): 12
 Operator: Milota
 Date: 3-26-04
 Time now: 5:40

AMBIENT DATA

Altimeter setting: 29.93 inHg

Laboratory temperature: 23 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>408</u>
mid	<u>414</u>	<u>416</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.571 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 1 L/min [TFR x (1 - DR)]

sample flow rate (TSFR) 2 L/min [TFR x DR]

Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.58 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>555</u>	<u>0.613</u>	<u>0.616</u>	<u>0.5</u>

START TIME: 5:53

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 1

Time now: 9:02

Run (sample): 11

AMBIENT DATA

Airport pressure: 2996 inHg

Laboratory temperature: 22 °C

END TIME: 9:02

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>556</u>	<u>557</u>

Sample flow rate (SFR): 1,573 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2,562 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0,614 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>904</u>	<u>904</u>	875 to 935	<u>208</u>
mid	<u>414</u>	<u>416</u>	379 to 437	none
zero	<u>0</u>	<u>0,5</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}): 0,615 [Span_{Diluted} / Span]

Dilution ratio difference: 0,16 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 9:05

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton 1 2

Wet-bulb temperature: 66

Run (sample): 12

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-26-04

Target Dilution Ratio (TDR): 0.6

Time now: 9:06

AMBIENT DATA

Altimeter setting: 2996 inHg

Laboratory temperature: 22 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	0 (0)	0.5	does not apply	480
span	905 (1527)	906	does not apply	206
mid	414	414	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.561 L/min

[1, 2, 3 = off; 4=meter]

Target

dilution flow rate (TDFR) _____ L/min

[TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min

[TFR x DR]

Set and read dilution meter: 2 scfh

[scfh = L/min * 2.12]

Sample flow rate (SFR): 1.586 L/min

[1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	555		0.613	

START TIME: 9:12

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

H₂ = 1650
PSI

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milora

Event (kiln charge): Hampton #2

Time now: 12

Run (sample): 12

AMBIENT DATA

Airport pressure: 29.93 inHg

Laboratory temperature: 21 °C

END TIME: 12:18

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>552</u>	<u>553</u>

Sample flow rate (SFR) : 1,576 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2,560 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.615 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>903</u>	<u>906</u>	875 to 935	<u>206</u>
mid	<u>413</u>	<u>413</u>	379 to 437	none
zero	<u>0</u>	<u>0</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}): 0.611 [Span_{Diluted} / Span]

Dilution ratio difference: 0.65 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 12:25

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton 1 2

Wet-bulb temperature: 66

Run (sample): 13

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-26-04

Target Dilution Ratio (TDR): 0.6

Time now: 12:25

AMBIENT DATA

Altimeter setting: 29.93 inHg

Laboratory temperature: 21 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>905</u>	does not apply	<u>202</u>
mid	<u>414</u>	<u>414</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2,551 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1,590 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>551</u>	<u>0.620</u>	<u>0.623</u>	<u>0.48</u>

START TIME: 12:30

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milora

Event (kiln charge): Hampton #2

Time now: 5:00

Run (sample): 12

AMBIENT DATA

Airport pressure: 29.88 inHg

Laboratory temperature: 21 °C

END TIME: 5:00

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>552</u>	<u>555</u>

Sample flow rate (SFR) : 1.584 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: _____ scfh _____ L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2.558 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.619 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>900</u>	<u>901</u>	875 to 935	
mid	<u>411</u>	<u>412</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	

Dilution ratio (DR_{Span}): 0.613 [Span_{Diluted} / Span]

Dilution ratio difference: 0.98 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 5:05

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton 1 2

Wet-bulb temperature: 66

Run (sample): 14

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-26-04

Target Dilution Ratio (TDR): 0.6

Time now: 5:00

AMBIENT DATA

Altimeter setting: 29.88 inHg

Laboratory temperature: 21 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>905</u>	does not apply	<u>212</u>
mid	<u>414</u>	<u>415</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.551 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.561 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>552</u>	<u>0.611</u>	<u>0.612</u>	<u>0.16</u>

START TIME: 5:08

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milora

Event (kiln charge): Hampton #2

Time now: 9:47p

Run (sample): 14

AMBIENT DATA

Airport pressure: 29.90 inHg

Laboratory temperature: 21 °C

END TIME: 10:00

→ weighed bonds & restarted

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>551</u>	<u>552</u>

Sample flow rate (SFR): 1.569 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2.564 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.612

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>905</u>	<u>906</u>	875 to 935	<u>212</u>
mid	<u>414</u>	<u>414</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>480</u>

Dilution ratio (DR_{Span}): 0.609

[Span_{Diluted} / Span]

Dilution ratio difference: 0.5 %

[100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 10:12

Comments:

Charge: 3	Date		Time
Hampton Affiliates	Start: 3-29-04	9:15a	
Page: 1	End:		

Clock time	Run time	Run	Temperatures							Flows					Lir
			Box	Valve	Dry-bulb	Wet-bulb	Line	Analzr	Chiller	Flow 1	Flow 2	Flow 3	Dilution	F/M	
	hrs	#	°C	°C	°C	°C	°F	°C	°C	L/min	L/min	SCFH	SCFM	mL/min	int
9:13	Pre	-	126	145	-	-	223	-	-	0	0	0	-	-	
9:30	0:15	✓	125	145	52	36	223	-	-	5	0	0	-	-	
10:44	1:29	✓	125	145	71	61	223	-	-	200	0	0	-	-	
10:46	1:34	✓	125	145	71	61	223	-	-	167	22	100	-	-	
11:33	2:18	✓	125	145	75	62	223	-	-	166	26	100	2.6	-	
11:34	2:19	2	125	145	73	63	223	-	-	149	43	120	2.7	-	
1:40	4:25	✓	125	145	77	63	223	-	-	150	41	120	2.7	-	
2:36	5:21	✓	125	145	81	64	223	-	-	151	40	120	2.7	-	
4:21	7:04	3	125	145	83	66	223	-	-	155	50	160	2.7	-	
4:22	7:05	3	125	145	84	67	223	-	-	25	41	50	2.7	-	
5:16	8:01	3	125	145	85	69	223	-	-	85	39	40	2.7	-	
5:17	8:02	3	125	145	86	70	223	-	-	86	40	0	2.7	-	
9:44	12:30	4	125	145	95	80	223	-	-	35	41	0	2.7	-	
5:18A	20:05	5	125	145	101	74	223	-	-	80	0	0	3.7	-	
5:40	20:22	6	125	145	101	74	223	-	-	87	0	0	3.2	-	
7:36	22:02	6	125	145	101	72	223	-	-	76	0	0	3.2	-	
8:52	23:36	7	125	145	102	71	223	-	-	81	0	0	3	-	

0
100
100
120

16
50
3-30-04

AT 6:00^{run time} Flow 1 began to decrease

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 155-180

Event (kiln charge): Hampton 3

Wet-bulb temperature: 120-150

Run (sample): 1

Absolute humidity: _____

Operator: Milota

Percent moisture: 11-12%

Date: 3-29-04

Target Dilution Ratio (TDR): 14-12%

Time now: 9:08

AMBIENT DATA

Altimeter setting: 29.86 inHg

Laboratory temperature: 23 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>-</u>	does not apply	<u>482</u>
span	<u>905</u> (1527)	<u>-</u>	does not apply	<u>211</u>
mid	<u>415</u>	<u>-</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): _____ L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: _____ scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): _____ L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % $100 * (DR_{Span} - DR_{Flow}) / DR_{Flow}$
Span _{Diluted}	_____	_____	_____	_____

START TIME: 9:15

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 8:08 11:37

Run (sample): 1

AMBIENT DATA

Airport pressure: 29.85 inHg

Laboratory temperature: 23 °C

END TIME: 11:38

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u> </u>	<u> </u>

Sample flow rate (SFR) : L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>905</u>	<u>906</u>	875 to 935	<u>211</u>
mid	<u>414</u>	<u>413</u>	379 to 437	none
zero	<u>01</u>	<u>1</u>	-45 to +45	<u>482</u>

Dilution ratio (DR_{Span}): [Span_{Diluted} / Span]

Dilution ratio difference: % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check:

Comments: 3-29-04 8:54 Leak Check 19.2" Hg → 18.6" Hg @ 9:02

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 70+

Event (kiln charge): Hampton 3

Wet-bulb temperature: 60+

Run (sample): 2

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-29-04

Target Dilution Ratio (TDR): 0.85

Time now: 11:38

AMBIENT DATA

Altimeter setting: 29.85 inHg

Laboratory temperature: _____ °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	0 (0)	1	does not apply	482
span	905 (1527)	906	does not apply	211
mid	414	413	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.582 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 1.291 L/min [TFR x (1 - DR)]

sample flow rate (TSFR) 1.291 L/min [TFR x DR]

Set and read dilution meter: 2.7 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.316 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow}) / DR _{Flow}
Span _{Diluted}	462	0.511	0.510	0.2

START TIME: 11:43

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 2:35

Run (sample): 2

AMBIENT DATA

Airport pressure: 29.80 inHg

Laboratory temperature: 30 °C

END TIME: 2:37 p

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>452</u>	<u>453</u>

Sample flow rate (SFR): 1,324 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 1.74 scfh L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2,612 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.507

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>901</u>	<u>901</u>	875 to 935	
mid	<u>411</u>	<u>412</u>	379 to 437	none
zero	<u>0</u>	<u>0</u>	-45 to +45	

Dilution ratio (DR_{Span}): 0.502

[Span_{Diluted} / Span]

Dilution ratio difference: 1% %

[100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 2:43

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82 →

Event (kiln charge): Hampton 3

Wet-bulb temperature: 64 →

Run (sample): 3

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-29-04

Target Dilution Ratio (TDR): 0.5

Time now: 2:38

AMBIENT DATA

Altimeter setting: 29.80 inHg

Laboratory temperature: 30 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0</u>	does not apply	<u>482</u>
span	<u>905</u> (1527)	<u>904</u>	does not apply	<u>213</u>
mid	<u>413</u>	<u>414</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.628 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2.7 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.324 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow}) / DR _{Flow}
Span _{Diluted}	<u>454</u>	<u>0.502</u>	<u>0.504</u>	<u>0.4</u>

START TIME: 2:47

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 5:16

Run (sample): 3

AMBIENT DATA

Airport pressure: 29.93 inHg

Laboratory temperature: 26 °C

END TIME: 5:18

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>461</u>	<u>460</u>

Sample flow rate (SFR) : 1.336 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2.7 scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2.619 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.510 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>906</u>	<u>907</u>	875 to 935	<u>213</u>
mid	<u>415</u>	<u>416</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>482</u>

Dilution ratio (DR_{Span}): 0.501 [Span_{Diluted} / Span]

Dilution ratio difference: 0.2 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 5:22

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: _____

Run (sample): 11

AMBIENT DATA

Airport pressure: 3020 inHg

Laboratory temperature: 21 °C

END TIME: 6:55a

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>491</u>	<u> </u>

Sample flow rate (SFR) : 1,408 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 27 scfh L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2,578 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.546

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>905</u>	<u> </u>	875 to 935	<u>208</u>
mid	<u>444</u>	<u> </u>	379 to 437	<u>none</u>
zero	<u>0</u>	<u> </u>	-45 to +45	<u>482</u>

Dilution ratio (DR_{Span}): 0.543

[Span_{Diluted} / Span]

Dilution ratio difference: 0.5 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 7:03

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: _____

Event (kiln charge): Hampton 3

Wet-bulb temperature: _____

Run (sample): 11

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-30-04

Target Dilution Ratio (TDR): 0.55

Time now: 9:50

AMBIENT DATA

Altimeter setting: 30.22 inHg

Laboratory temperature: 22 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	
span	<u>905</u> (1527)	<u>906</u>	does not apply	
mid	<u>414</u>	<u>416</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.594 L/min

[1, 2, 3 = off; 4=meter]

Target

dilution flow rate (TDFR) 1.19 L/min

[TFR x (1 - DR)]

sample flow rate (TSFR) 1.4 L/min

[TFR x DR]

Set and read dilution meter: 2.53 scfh

[scfh = L/min * 2.12]

Sample flow rate (SFR): 1.390 L/min

[1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>490</u>	<u>0.541</u>	<u>0.536</u>	<u>1.03</u>

START TIME: 10:22

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 9:40

Run (sample): 10

AMBIENT DATA

Airport pressure: 30.22 inHg

Laboratory temperature: 22 °C

END TIME: 9:45

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>446</u>	<u>445</u>

Sample flow rate (SFR) : 1.289 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 375 scfh 2.585 L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 3.874 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.499

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>905</u>	<u>905</u>	875 to 935	<u>208</u>
mid	<u>415</u>	<u>416</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>482</u>

Dilution ratio (DR_{Span}): 0.493

[Span_{Diluted} / Span]

Dilution ratio difference: 1.2 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 10:16 p

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: _____

Event (kiln charge): Hampton 3

Wet-bulb temperature: 68 ↓

Run (sample): 10

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-30-04

Target Dilution Ratio (TDR): 0.5

Time now: 5:00

AMBIENT DATA

Altimeter setting: 30.8 inHg

Laboratory temperature: 22 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>482</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>208</u>
mid	<u>415</u>	<u>414</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2573 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) _____ L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 2.7 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.269 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow}) / DR _{Flow}
Span _{Diluted}	<u>448</u>	<u>0.495</u>	<u>0.493</u>	<u>0.3%</u>

START TIME: 5:14 P

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: _____

Run (sample): 89

AMBIENT DATA

Airport pressure: 30.18 inHg

Laboratory temperature: 22 °C

END TIME: _____

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>446</u>	<u>446</u>

Sample flow rate (SFR) : 1.267 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2.7 scfh _____ L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2.567 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.494 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>904</u>	<u>903</u>	875 to 935	
mid	<u>413</u>	<u>413</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	

Dilution ratio (DR_{Span}): 0.493 [Span_{Diluted} / Span]

Dilution ratio difference: 0.04 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 5:12

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: _____

Event (kiln charge): Hampton 3

Wet-bulb temperature: _____

Run (sample): 89

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: _____

Target Dilution Ratio (TDR): _____

Time now : _____

AMBIENT DATA

Altimeter setting: 3013 inHg

Laboratory temperature: 23 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0</u>	does not apply	<u>482</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>206</u>
mid	<u>412</u>	<u>413</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.578 L/min

[1, 2, 3 = off; 4=meter]

Target

dilution flow rate (TDFR) _____ L/min

[TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min

[TFR x DR]

Set and read dilution meter: 2.7 scfh

[scfh = L/min * 2.12]

Sample flow rate (SFR): 1.291 L/min

[1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow}) / DR _{Flow}
Span _{Diluted}	<u>452</u>	<u>0.499</u>	<u>0.501</u>	

START TIME: 3pm

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: _____

Run (sample): 8

AMBIENT DATA

Airport pressure: _____ inHg

Laboratory temperature: _____ °C

END TIME: _____

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>435</u>	<u>436</u>

Sample flow rate (SFR) : 1245 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2.8 scfh _____ L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2.571 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.48

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>903</u>	<u>904</u>	875 to 935	<u>205</u>
mid	<u>413</u>	<u>414</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>481</u>

Dilution ratio (DR_{Span}): 0.48

[Span_{Diluted} / Span]

Dilution ratio difference: 0 %

[100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 2:57

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 102

Event (kiln charge): Hampton 3

Wet-bulb temperature: 70

Run (sample): 8

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-30-04

Target Dilution Ratio (TDR): 0.5

Time now: 1145

AMBIENT DATA

Altimeter setting: 30.13 inHg

Laboratory temperature: 23 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0.5</u>	does not apply	<u>281</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>205</u>
mid	<u>415</u>	<u>415</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 3,565 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 1,283 L/min [TFR x (1 - DR)]

sample flow rate (TSFR) 1,283 L/min [TFR x DR]

Set and read dilution meter: 2.7 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1,255 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>443</u>	<u>0.49</u>	<u>0.49</u>	<u>0</u>

START TIME: 12:10

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 11:48

Run (sample): 7

AMBIENT DATA

Airport pressure: 3013 inHg

Laboratory temperature: 23 °C

END TIME: 11:48

CHECK DILUTION FLOW AFTER RUN 396

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>400</u>	<u>400.399</u>

Sample flow rate (SFR): 1.135 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 3 scfh L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2.563 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.443

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>911</u>	<u>911</u>	875 to 935	<u>210</u>
mid	<u>412</u>	<u>412</u>	379 to 437	none
zero	<u>0</u>	<u>0.15</u>	-45 to +45	<u>282</u>

Dilution ratio (DR_{Span}): 0.435

[Span_{Diluted} / Span]

Dilution ratio difference: 1.8 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 12:05

Comments:

Maybe some unreliable data in this run
Computer did not agree with analyzer & actually
during post check

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 105

Event (kiln charge): Hampton 3

Wet-bulb temperature: 72 ↓

Run (sample): 7

Absolute humidity: 0.31

Operator: Milota

Percent moisture: 33

Date: 8:30

Target Dilution Ratio (TDR): 0.45

Time now: 3-30-04

AMBIENT DATA

Altimeter setting: 30.04 inHg

Laboratory temperature: 23 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>482</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>210</u>
mid	<u>415</u>	<u>416</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2,570 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 1,413 L/min [TFR x (1 - DR)]

sample flow rate (TSFR) 1,156 L/min [TFR x DR]

Set and read dilution meter: 3 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1,131 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>400</u>	<u>0.44</u>	<u>0.44</u>	<u>0</u>

START TIME: 8:47

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 8:30

Run (sample): 6

AMBIENT DATA

Airport pressure: 30.8 inHg

Laboratory temperature: 23 °C

END TIME: 8:40

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>33 342</u>	<u>343</u>

Sample flow rate (SFR) : 0.975 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 3.3 scfh 1.858 L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2.571 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): 0.379

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>903</u>	<u>904</u>	875 to 935	<u>210</u>
mid	<u>413</u>	<u>414</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>482</u>

Dilution ratio (DR_{Span}): 0.379

[Span_{Diluted} / Span]

Dilution ratio difference:

0 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check:

8:44

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 100-105

Event (kiln charge): Hampton 3

Wet-bulb temperature: 74 ↓

Run (sample): 6

Absolute humidity: 0.37

Operator: Milota

Percent moisture: 37

Date: 3-30-04

Target Dilution Ratio (TDR): 0.4

Time now: 5:19

AMBIENT DATA

Altimeter setting: 3001 inHg

Laboratory temperature: 25 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> ⁽⁰⁾	<u>1</u>	does not apply	<u>882</u>
span	<u>905</u> ⁽¹⁵²⁷⁾	<u>906</u>	does not apply	<u>210</u>
mid	<u>414</u>	<u>413</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2,589 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 1.0 L/min 1.51 [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 222 scfh 3.2 [scfh = L/min * 2.12]

Sample flow rate (SFR): 1,504 L/min 1.005 [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} /Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>352</u>	<u>0.389</u>	<u>0.388</u>	<u>0.2</u>

START TIME: 5:34A

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 5:18 a

Run (sample): 12 5

AMBIENT DATA

Airport pressure: 3001 inHg

Laboratory temperature: 25 °C

END TIME: 5:21a 3-30-04

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>239</u>	<u>238</u>

Sample flow rate (SFR) : 707 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 39 scfh 1.84 L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2,588 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}):

0.273

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>904</u>	<u>904</u>	875 to 935	<u>208</u>
mid	<u>413</u>	<u>412</u>	379 to 437	none
zero	<u>0</u>	<u>1</u>	-45 to +45	<u>482</u>

Dilution ratio (DR_{Span}):

0.264

[Span_{Diluted} / Span]

Dilution ratio difference:

34 %

[100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check:

5:27

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 206-210 F

Event (kiln charge): Hampton 3

Wet-bulb temperature: 80 c

Run (sample): 5

Absolute humidity: 0.65

Operator: Milota

Percent moisture: 46

Date: 3-29-04

Target Dilution Ratio (TDR): 0.32

Time now: 9:45

AMBIENT DATA

Altimeter setting: 29.96 inHg

Laboratory temperature: _____ °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> ⁽⁰⁾	<u>1</u>	does not apply	<u>482</u>
span	<u>905</u> ⁽¹⁵²⁷⁾	<u>906</u>	does not apply	<u>208</u>
mid	<u>415</u>	<u>416</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2,620 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 1,78 L/min [TFR x (1 - DR)]

sample flow rate (TSFR) 0,838 L/min [TFR x DR]

Set and read dilution meter: 3.8 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 7423 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>258</u>	<u>0,285</u>	<u>0,283</u>	

START TIME: 9:58pm

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milota

Event (kiln charge): Hampton 3

Time now: 9:46 P

Run (sample): 4

AMBIENT DATA

Airport pressure: 29.96 inHg

Laboratory temperature: 27 °C

END TIME: 9:48

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	338	342

Sample flow rate (SFR) : 0.980 L/min [1= on, 2, 3 = off, 4=meter]

Read dilution meter: 3.9 scfh L/min [L/min = scfh*0.472]

Total flow rate (TFR): 2.614 L/min [1, 2, 3 = off; 4=meter]
(attach print out with all four sets of data)

Dilution ratio (DR_{Flow}): 0.375 [SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	912	912	875 to 935	212
mid	416	415	379 to 437	none
zero	0	1	-45 to +45	482

Dilution ratio (DR_{Span}): 0.371 [Span_{Diluted} / Span]

Dilution ratio difference: 1.1 % [100*(Abs(DR_{Span} - DR_{Flow}))/DR_{Flow}]

End time for check: 9:55 P

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 87 →

Event (kiln charge): Hampton 3

Wet-bulb temperature: 70 →

Run (sample): 4

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-29

Target Dilution Ratio (TDR): 0.4

Time now: 5:15 p

AMBIENT DATA

Altimeter setting: 29.93 inHg

Laboratory temperature: 26 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>1</u>	does not apply	<u>482</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>212</u>
mid	<u>415</u>	<u>415</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.617 L/min [1, 2, 3 = off; 4=meter]

Target dilution flow rate (TDFR) 1.5702 L/min [TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min [TFR x DR]

Set and read dilution meter: 3.3 scfh [scfh = L/min * 2.12]

Sample flow rate (SFR): 1.016 L/min [1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>348</u>	<u>0.385</u>	<u>0.388</u>	<u>0.8</u>

START TIME: 5:25

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - POST

Operator: Milora

Event (kiln charge): Hampton #2

Time now: 10:10

Run (sample): _____

AMBIENT DATA

Airport pressure: 29.1 inHg

Laboratory temperature: _____ °C

END TIME: 2 7AM

CHECK DILUTION FLOW AFTER RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	Computer
Span _{Diluted}	<u>557</u>	<u> </u>

Sample flow rate (SFR) : 1,582 L/min

[1= on, 2, 3 = off, 4=meter]

Read dilution meter: 2 scfh _____ L/min

[L/min = scfh*0.472]

Total flow rate (TFR): 2,575 L/min
(attach print out with all four sets of data)

[1, 2, 3 = off; 4=meter]

Dilution ratio (DR_{Flow}): _____

[SFR / TFR]

CHECK OF ANALYZER CALIBRATION

[1, 2=off; 3=on, 4=vent]

	Analyzer	Computer	Within range	Pot settings
span	<u>920</u>		875 to 935	
mid	<u>423</u>		379 to 437	none
zero			-45 to +45	

Dilution ratio (DR_{Span}): _____

[$Span_{Diluted} / Span$]

Dilution ratio difference: _____ %

[$100 * (Abs(DR_{Span} - DR_{Flow})) / DR_{Flow}$]

End time for check: _____

Comments:

FIELD DATA SHEET FOR TOTAL HYDROCARBON ANALYZER - PRE

BACKGROUND INFORMATION

Dry-bulb temperature: 82

Event (kiln charge): Hampton 1 2

Wet-bulb temperature: 46

Run (sample): 15

Absolute humidity: _____

Operator: Milota

Percent moisture: _____

Date: 3-26-04

Target Dilution Ratio (TDR): 0.6

Time now: 10:16

AMBIENT DATA

Altimeter setting: 29.90 inHg

Laboratory temperature: 21 °C

ANALYZER CALIBRATION

[1, 2 = off; 3=on; 4=vent]

	Analyzer, ppm	Computer	Within range	Pot settings
zero	<u>0</u> (0)	<u>0</u>	does not apply	<u>480</u>
span	<u>905</u> (1527)	<u>906</u>	does not apply	<u>212</u>
mid	<u>414</u>	<u>415</u>	379 to 437	none

SET DILUTION FLOW BEFORE RUN

Total flow rate (TFR): 2.565 L/min

[1, 2, 3 = off; 4=meter]

Target

dilution flow rate (TDFR) _____ L/min

[TFR x (1 - DR)]

sample flow rate (TSFR) _____ L/min

[TFR x DR]

Set and read dilution meter: 2 scfh

[scfh = L/min * 2.12]

Sample flow rate (SFR): 1.570 L/min

[1 = on; 2, 3 = off; 4=meter]

CHECK DILUTION FLOW BEFORE RUN

[1, 3=on; 2=off; 4=vent]

	Analyzer	DR _{Span} [Span _{Diluted} / Span]	DR _{Flow} [SFR / TFR]	Difference, % 100*(DR _{Span} - DR _{Flow})/DR _{Flow}
Span _{Diluted}	<u>555</u>	<u>0.613</u>	<u>0.612</u>	<u>0.16</u>

START TIME: 10:15

[1, 2, 5 = on; 3, 4 = off; tank valves off]

ANALYZER RANGE: 2

[60 < computer reading < 750]

Appendix 4. Calibration Data

CERTIFICATE OF ANALYSIS
EPA Protocol Gas**CUSTOMER**INDUSTRIAL WELDING SUPPLY
3415 S PACIFIC BLVD
ALBANY, OR 973210000

CUSTOMER PO NO: 40734

Previous Certification Date(s):

CYLINDER NO : CC62620
EXPIRATION DATE : 11-Jan-2007
CERTIFICATION DATE : 12-Jan-2004
CYLINDER PRESSURE : 2000 psig
PRODUCT ID NO : 24011968
LOT NUMBER : 592976**ANALYTICAL INFORMATION**

This calibration standard has been certified per the 1997 EPA Traceability Protocol, Document EPA-600/97/121, Using Procedure G1. All Values certified to be +/-1% NIST Traceable.

Do Not Use This Cylinder below 150 psig. i.e. 1.0 Megapascal

Analytical Results

Components	Requested Mixture	Certified Concentration	Analytical Uncertainty	Assay Dates
PROPANE	900.00 ppm	905 ppm	+/-1.00% NIST Traceable	01/12/04
IR	BALANCE GAS			

CALIBRATION STANDARDS USED IN ASSAY

Type	LOT ID	Cylinder No	Concentration	Expiration
NTRM 81669	99060605	XC003606B	453.00 +/- 4.00 ppm C3H8/AIR	06/01/07

ANALYTICAL INSTRUMENTS USED IN ASSAY

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
H-P 5890 3022A29265	Gas Chromatography	01/12/04

CERTIFICATE OF ANALYSIS**To :**INDUSTRIAL WELDING SUPPLY
3415 PACIFIC BLVD SW
ALBANY, OR 973217725**Material Submitted** : C3H8 400M (AIR) CERT 152
Specification No. : 24017070**Cylinder Number** : CC5263
Cyl.Size/Valve : 152/590 BR 3360**Customer P.O.** : 40734
Order Number : 0102500677
Date Reported : 13-JAN-2004
Fill Date : 13-JAN-2004
Expired Date : 12-JAN-2007
Lot Number : 593083**Pressure** : 2000 psig

Component	Specifications	Min	Max	Concentration
PROPANE	400 ppm	360	440	412 ppm
AIR				BALANCE

BOC GASES
PORT ALLEN, LOUISIANA

Cylinder # : CC5263
CU # : 0102500577
Pressure : 2000 psig
CGA Outlet : 152/590 BR 3360
Fill : 13-JAN-2004
Expired : 12-JAN-2007
Material : C3H8 400M (AIR) CERT 152

CYLINDER CONTENT ANALYSIS

Component	Concentration
propane	412 ppm
air	BALANCE

BOC GASES
VANCOUVER, WASHINGTON

Order Number : GS-1504
Approx. Pressure : 2200 psig
CGA Outlet : CGA348
Fill Date : 28-APR-2003
Expiration Date : 28-APR-2006
Lot Number : 25902

Air, Zero 0.1

CYLINDER CONTENT ANALYSIS	
Component	Concentration
Oxygen	20.6 %
Nitrogen	Balance
Moisture	< 3.0 ppm
Total Hydrocarbon	< 0.1 ppm

TC Calibration	
6/10/2002	
Omega Calibrator C	PC Readout C
30.0	30.0
50.0	50.1
70.0	70.1
90.0	90.1
110.0	110.0



Flow
Calibration Record Sheet
(200 SLM)

ERA #: 128989W

Customer: OREGON STATE UNIVERSITY

MKS Transfer Standard Type: 1559A-200L-SV

Serial Number: WS 136

MKS Primary Standard Type: A-200-1

Serial Number: 14952-1-1

Standard Flow Rate (SLM)	UUT Flow Rate (SLM)	UUT Error (SLM)	Percent of full scale Error
0.00	0.000	0.000	0.000%
50.000	50.880	0.880	0.440%
100.000	99.880	-0.120	-0.060%
150.000	150.040	0.040	0.020%
200.000	200.000	0.000	0.000%

UUT Model: 1559A-200L-SV

UUT Process Gas: N2

Process Gas used: N2

Date of Calibration: 05/10/00

UUT Serial #: 000317785

UUT Range: 200 SLM

Calibrated by: DP

Verified by:

Notes:

1. All units must be operated on regulated heat (Power on) for a minimum of one hour before any adjustment is made.
2. Flowmeters and/or Controllers are Calibrated at atmospheric pressure.
3. This Calibration is referenced to 0 Degrees Centigrade and 760 Torr.

3350 Scott Blvd., Bldg. #4, Santa Clara, CA 95054

This MKS Certificate or report shall not be reproduced except in full, without the written approval of the Laboratory (MKS).